

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2930

Name: RITTER 1  
Sterilizer ID: V2178198  
02/22/2023 01:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.7 F 0.0  
2:00 181.2 F 4.3  
4:00 193.0 F 9.0  
6:00 212.3 F 7.7  
8:00 230.0 F 8.2  
10:00 241.8 F 13.4  
12:00 239.2 F 10.6  
14:00 236.7 F 8.7  
16:00 251.2 F 15.8  
18:00 263.4 F 22.8  
20:00 271.3 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 28.9  
2:00 271.2 F 28.2  
3:00 272.5 F 29.0  
4:00 271.8 F 28.5  
  
Min 271.2 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.7 F 0.0  
5:00 234.3 F 0.0  
10:00 228.3 F 0.0  
15:00 223.0 F 0.0  
20:00 229.9 F 0.0  
25:00 228.7 F 0.0  
30:00 231.8 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

02/22/2023 02:11 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_