
Midmark M11 - v1.0.0
Total Cycles: 610

Name: MIDMARK
Sterilizer ID: V2178198
01/20/2021 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.8 F 0.0
2:00 72.5 F 0.0
4:00 86.8 F 0.1
6:00 122.3 F 0.1
8:00 158.8 F 0.3
10:00 186.1 F 1.9
12:00 202.1 F 5.5
14:00 217.4 F 9.1
16:00 229.1 F 13.2
18:00 230.1 F 8.0
20:00 241.4 F 12.9
22:00 239.0 F 10.3
24:00 235.4 F 8.2
26:00 250.6 F 15.5
28:00 263.1 F 22.7
30:00 271.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.1 F 28.4
2:00 272.7 F 28.9
3:00 272.0 F 28.7
4:00 271.8 F 28.2

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.6 F 0.0
5:00 221.2 F 0.0
10:00 218.1 F 0.0
15:00 226.0 F 0.0
20:00 231.0 F 0.0
25:00 224.2 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:06
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:07:53
CYCLE COMPLETE

01/20/2021 08:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____