

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2888

Name: RITTER2  
Sterilizer ID: V2313228  
09/15/2023 08:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.0 F 0.0  
2:00 70.2 F 0.0  
4:00 81.3 F 0.0  
6:00 118.1 F 0.0  
8:00 152.2 F 0.0  
10:00 172.8 F 2.7  
12:00 193.5 F 6.1  
14:00 222.7 F 9.6  
16:00 229.6 F 13.3  
18:00 227.8 F 7.3  
20:00 238.5 F 11.5  
22:00 233.2 F 8.1  
24:00 245.0 F 12.9  
26:00 243.5 F 12.2  
28:00 255.2 F 18.2  
30:00 265.4 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.6  
1:00 272.5 F 29.1  
2:00 271.6 F 29.1  
3:00 271.9 F 28.8  
4:00 272.1 F 29.4

Min 271.1 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
5:00 234.9 F 0.0  
10:00 224.2 F 0.0  
15:00 183.4 F 0.0  
20:00 176.9 F 0.0  
25:00 175.1 F 0.0  
30:00 174.2 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:52  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:10:03  
CYCLE COMPLETE

09/15/2023 09:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_