
Midmark M11 - v1.0.0
Total Cycles: 1089

Name: RITTER 1
Sterilizer ID: V2178198
08/12/2021 07:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.3 F 0.0
2:00 72.4 F 0.0
4:00 79.2 F 0.0
6:00 123.6 F 0.1
8:00 155.9 F 0.1
10:00 175.5 F 0.2
12:00 199.4 F 1.4
14:00 212.7 F 4.5
16:00 221.2 F 7.0
18:00 229.8 F 10.1
20:00 234.6 F 10.0
22:00 238.2 F 10.8
24:00 234.2 F 8.0
26:00 242.3 F 11.0
28:00 246.2 F 13.3
30:00 258.3 F 19.7
32:00 269.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.2 F 28.4
2:00 272.6 F 28.9
3:00 271.6 F 28.1
4:00 272.5 F 28.7

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 186.0 F 0.0
10:00 185.7 F 0.0
15:00 193.1 F 0.0
20:00 188.5 F 0.0
25:00 182.0 F 0.0
30:00 181.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 33:05
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:11:04
CYCLE COMPLETE

08/12/2021 08:27 AM
ACCEPTED: X
REJECTED:
Operator: U02