
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001287
10:16 07/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:50
STERILIZING 5:03
136.0 C 221kPa 5:03
136.0 C 222kPa 5:33
136.4 C 222kPa 6:03
136.4 C 222kPa 6:33
136.1 C 223kPa 7:03
136.1 C 223kPa 7:33
136.1 C 223kPa 8:03
136.4 C 225kPa 8:33
136.4 C 225kPa 8:33
VENTING 8:33
AIR DRYING 9:25
CYCLE COMPLETE 24:25

Digital Signature #
906CAA36F39F7CAA
