

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002762  
15:21 30/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:11  
PRESSURIZING 3:12  
STERILIZING 4:18  
136.4 C 222kPa 4:18  
136.4 C 224kPa 4:48  
136.4 C 224kPa 5:18  
136.4 C 224kPa 5:48  
136.2 C 221kPa 6:18  
136.4 C 224kPa 6:48  
136.4 C 225kPa 7:18  
136.1 C 222kPa 7:48  
136.1 C 222kPa 7:48  
VENTING 7:48  
AIR DRYING 8:44  
CYCLE COMPLETE 9:10

-----  
Digital Signature #  
F9BEBBE035C08C95  
-----