
Midmark M11 - v1.0.5
Total Cycles: 4394

Name: RITTER2
Sterilizer ID: V2313228
03/03/2025 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 70.5 F 0.2
4:00 86.0 F 0.9
6:00 100.5 F 3.3
8:00 145.2 F 7.0
10:00 173.7 F 11.6
12:00 200.5 F 2.9
14:00 214.1 F 6.5
16:00 225.8 F 10.4
18:00 229.1 F 8.0
20:00 239.2 F 11.5
22:00 237.1 F 9.4
24:00 250.9 F 16.1
26:00 262.9 F 23.2
28:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.4 F 29.0
2:00 272.9 F 29.4
3:00 272.9 F 29.8
4:00 272.0 F 29.1

Min 271.7 F 28.8
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 189.1 F 0.0
10:00 192.5 F 0.0
15:00 196.8 F 0.0
20:00 204.4 F 0.0
25:00 211.2 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:06
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:06:09
CYCLE COMPLETE

03/03/2025 11:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____