

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3040

Name: RITTER2  
Sterilizer ID: V2313228  
10/23/2023 08:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.5 F 0.0  
2:00 71.7 F 0.0  
4:00 83.4 F 0.4  
6:00 94.3 F 2.3  
8:00 126.7 F 5.2  
10:00 161.5 F 8.7  
12:00 186.5 F 12.4  
14:00 206.6 F 3.9  
16:00 217.1 F 7.4  
18:00 227.9 F 11.2  
20:00 228.2 F 6.9  
22:00 239.4 F 11.9  
24:00 237.3 F 9.8  
26:00 250.5 F 16.1  
28:00 261.6 F 22.5  
30:00 272.0 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.1  
1:00 272.4 F 30.0  
2:00 271.7 F 29.1  
3:00 272.4 F 29.9  
4:00 272.2 F 29.4  
  
Min 271.3 F 29.0  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.3 F 0.0  
5:00 203.4 F 0.0  
10:00 195.6 F 0.0  
15:00 188.8 F 0.0  
20:00 184.8 F 0.0  
25:00 188.4 F 0.0  
30:00 191.2 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:23  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:08:22  
CYCLE COMPLETE

10/23/2023 09:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_