
Midmark M11 - v1.0.0
Total Cycles: 1341

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2021 01:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 111.9 F 0.0
2:00 113.8 F 0.1
4:00 128.2 F 0.1
6:00 168.9 F 0.3
8:00 204.9 F 3.8
10:00 216.8 F 7.3
12:00 225.3 F 10.9
14:00 229.8 F 6.6
16:00 235.2 F 10.0
18:00 240.3 F 9.4
20:00 244.3 F 12.7
22:00 242.6 F 11.3
24:00 255.0 F 17.6
26:00 265.5 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.0
1:00 272.0 F 28.6
2:00 272.2 F 28.4
3:00 271.4 F 28.1
4:00 272.6 F 28.8

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 214.2 F 0.0
10:00 217.8 F 0.0
15:00 218.7 F 0.0
20:00 223.1 F 0.0
25:00 222.7 F 0.0
30:00 212.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 27:42
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:06:10
CYCLE COMPLETE

11/08/2021 03:03 PM
ACCEPTED: X
REJECTED:
Operator: U01