
Midmark M11 - v1.0.0
Total Cycles: 4275

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2024 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.9 F 0.0
2:00 122.0 F 1.4
4:00 134.2 F 3.9
6:00 161.6 F 7.1
8:00 190.8 F 11.0
10:00 204.4 F 3.1
12:00 222.0 F 6.8
14:00 231.4 F 10.8
16:00 231.7 F 7.1
18:00 240.5 F 11.7
20:00 238.0 F 9.4
22:00 250.3 F 15.4
24:00 261.0 F 21.4
26:00 271.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.6 F 29.1
2:00 272.1 F 28.4
3:00 272.4 F 28.9
4:00 272.5 F 28.7

Min 271.6 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 215.5 F 0.0
10:00 234.7 F 0.0
15:00 230.1 F 0.0
20:00 227.9 F 0.0
25:00 215.7 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:05:08
CYCLE COMPLETE

02/06/2024 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____