

-----  
Midmark M11 - v1.0.5  
Total Cycles: 398

Name: RITTER2  
Sterilizer ID: V2313228  
09/28/2021 06:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.0 F 0.0  
2:00 73.4 F 0.0  
4:00 81.3 F 0.1  
6:00 114.1 F 0.1  
8:00 155.4 F 0.7  
10:00 196.4 F 4.0  
12:00 192.3 F 8.1  
14:00 207.8 F 11.9  
16:00 221.0 F 6.6  
18:00 235.0 F 10.9  
20:00 232.0 F 7.6  
22:00 243.2 F 12.8  
24:00 241.4 F 11.3  
26:00 253.5 F 17.5  
28:00 264.0 F 23.8  
30:00 272.1 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 272.1 F 29.3  
2:00 272.0 F 29.4  
3:00 271.9 F 29.3  
4:00 272.1 F 29.2

Min 271.8 F 29.2  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.6 F 0.0  
5:00 230.4 F 0.0  
10:00 201.3 F 0.0  
15:00 183.0 F 0.0  
20:00 173.5 F 0.0  
25:00 169.9 F 0.0  
30:00 170.2 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 30:47  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:08:42  
CYCLE COMPLETE

09/28/2021 07:28 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01