

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2171

Name: RITTER2  
Sterilizer ID: V2313228  
03/22/2023 04:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.6 F 0.0  
2:00 161.0 F 2.9  
4:00 174.9 F 6.5  
6:00 198.5 F 11.1  
8:00 216.0 F 5.2  
10:00 227.9 F 10.0  
12:00 232.6 F 8.6  
14:00 242.1 F 12.8  
16:00 242.3 F 11.7  
18:00 254.6 F 18.2  
20:00 265.5 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.8  
1:00 272.6 F 29.5  
2:00 271.9 F 29.2  
3:00 273.0 F 29.7  
4:00 271.8 F 29.6  
  
Min 271.0 F 28.7  
Max 273.2 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 206.2 F 0.0  
10:00 214.9 F 0.0  
15:00 219.4 F 0.0  
20:00 225.1 F 0.0  
25:00 229.8 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 21:44  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:00:36  
CYCLE COMPLETE

03/22/2023 05:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_