

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5479

Name: RITTER 1  
Sterilizer ID: V2178198  
04/04/2025 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.0 F 0.0  
2:00 135.8 F 1.7  
4:00 152.6 F 4.9  
6:00 176.4 F 8.4  
8:00 194.1 F 12.3  
10:00 215.4 F 3.6  
12:00 220.3 F 7.5  
14:00 231.0 F 11.6  
16:00 231.0 F 7.0  
18:00 242.7 F 12.1  
20:00 239.3 F 9.8  
22:00 250.9 F 15.5  
24:00 261.1 F 21.3  
26:00 270.9 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.0  
1:00 272.1 F 28.9  
2:00 272.5 F 28.8  
3:00 271.5 F 28.5  
4:00 272.6 F 29.2

Min 271.0 F 27.9  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.2 F 0.0  
5:00 205.3 F 0.0  
10:00 193.8 F 0.0  
15:00 192.6 F 0.0  
20:00 210.4 F 0.0  
25:00 219.2 F 0.0  
30:00 219.9 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 26:50  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:05:37  
CYCLE COMPLETE

04/04/2025 01:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_