
Midmark M11 - v1.0.5
Total Cycles: 1689

Name: RITTER2
Sterilizer ID: V2313228
11/07/2022 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.1 F 0.0
2:00 135.0 F 0.0
4:00 161.9 F 0.8
6:00 190.4 F 4.1
8:00 213.3 F 8.9
10:00 223.7 F 10.0
12:00 233.2 F 10.2
14:00 234.5 F 8.4
16:00 242.0 F 11.5
18:00 246.4 F 13.7
20:00 259.2 F 20.6
22:00 270.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.0 F 29.0
2:00 272.3 F 29.5
3:00 271.5 F 28.7
4:00 272.4 F 29.4

Min 271.3 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 215.1 F 0.0
10:00 237.5 F 0.0
15:00 229.2 F 0.0
20:00 221.7 F 0.0
25:00 217.9 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 22:48
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

11/07/2022 03:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____