
Midmark M11 - v1.0.5
Total Cycles: 3425

Name: RITTER2
Sterilizer ID: V2313228
03/01/2024 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.2 F 0.0
2:00 71.6 F 0.0
4:00 78.6 F 0.0
6:00 113.1 F 0.6
8:00 145.9 F 2.7
10:00 168.0 F 6.0
12:00 186.2 F 9.4
14:00 202.5 F 13.1
16:00 220.4 F 5.8
18:00 229.2 F 9.5
20:00 236.2 F 12.9
22:00 236.2 F 9.7
24:00 238.2 F 9.8
26:00 245.1 F 12.9
28:00 256.3 F 18.7
30:00 266.2 F 24.9
32:00 272.0 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.1
2:00 272.5 F 29.5
3:00 271.7 F 29.4
4:00 272.0 F 29.0

Min 271.4 F 28.9
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.3 F 0.0
5:00 234.7 F 0.0
10:00 215.3 F 0.0
15:00 201.5 F 0.0
20:00 176.1 F 0.0
25:00 174.3 F 0.0
30:00 175.7 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:30
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:10:33
CYCLE COMPLETE

03/01/2024 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____