
Midmark M11 - v1.0.5
Total Cycles: 4360

Name: RITTER2
Sterilizer ID: V2313228
02/19/2025 06:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.1 F 0.0
2:00 141.0 F 2.0
4:00 150.3 F 5.2
6:00 179.0 F 9.8
8:00 214.7 F 7.1
10:00 225.9 F 9.8
12:00 230.3 F 8.0
14:00 241.1 F 12.7
16:00 240.6 F 11.2
18:00 253.8 F 17.9
20:00 265.5 F 25.0
22:00 271.0 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.7 F 29.9
1:00 272.1 F 30.1
2:00 272.1 F 30.3
3:00 272.5 F 30.2
4:00 272.4 F 30.1

Min 271.7 F 29.8
Max 272.7 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 209.9 F 0.0
10:00 228.2 F 0.0
15:00 216.1 F 0.0
20:00 209.4 F 0.0
25:00 206.5 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 22:20
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

02/19/2025 07:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____