
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007805
13:49 28/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:28
STERILIZING 3:36
136.4 C 220kPa 3:36
136.5 C 224kPa 4:06
135.8 C 220kPa 4:36
135.8 C 220kPa 5:06
136.2 C 218kPa 5:36
136.4 C 223kPa 6:06
136.3 C 224kPa 6:36
136.4 C 224kPa 7:06
136.4 C 224kPa 7:06
VENTING 7:06
AIR DRYING 7:57
CYCLE COMPLETE 27:16

Digital Signature #
30B8F50E15D72105
