
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002654
10:20 07/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:14
STERILIZING 4:26
136.0 C 220kPa 4:26
136.2 C 223kPa 4:56
136.4 C 223kPa 5:26
136.3 C 224kPa 5:56
136.1 C 223kPa 6:26
136.4 C 226kPa 6:56
136.4 C 225kPa 7:26
136.4 C 225kPa 7:56
136.4 C 225kPa 7:56
VENTING 7:56
AIR DRYING 8:50
CYCLE COMPLETE 13:53

Digital Signature #
9C7202AAE021A522
