
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000409
12:32 26/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:22
STERILIZING 4:26
136.0 C 230kPa 4:26
136.2 C 230kPa 4:56
136.4 C 233kPa 5:26
136.2 C 231kPa 5:56
136.4 C 233kPa 6:26
136.4 C 233kPa 6:56
136.2 C 231kPa 7:26
136.2 C 230kPa 7:56
136.2 C 230kPa 7:56
VENTING 7:56
AIR DRYING 8:53
CYCLE COMPLETE 8:58

Digital Signature #
434A6503B4A3D361
