
Midmark M11 - v1.0.0
Total Cycles: 3588

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2023 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 77.0 F 0.0
4:00 98.2 F 0.0
6:00 115.3 F 0.0
8:00 148.2 F 0.9
10:00 169.4 F 2.5
12:00 181.5 F 4.7
14:00 195.5 F 7.5
16:00 207.6 F 10.5
18:00 219.6 F 10.8
20:00 224.2 F 6.9
22:00 233.6 F 10.8
24:00 233.1 F 8.0
26:00 242.9 F 12.2
28:00 240.0 F 10.3
30:00 251.3 F 15.8
32:00 261.7 F 21.7
34:00 271.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.8
2:00 271.5 F 28.0
3:00 272.6 F 28.9
4:00 271.8 F 28.2

Min 271.4 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.1 F 0.0
5:00 204.8 F 0.0
10:00 199.0 F 0.0
15:00 203.5 F 0.0
20:00 205.9 F 0.0
25:00 215.6 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 34:44
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:13:06
CYCLE COMPLETE

07/27/2023 09:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____