

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2487

Name: RITTER2  
Sterilizer ID: V2313228  
06/08/2023 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
2:00 193.4 F 4.0  
4:00 208.5 F 9.3  
6:00 221.3 F 7.7  
8:00 233.3 F 11.0  
10:00 234.5 F 8.9  
12:00 239.1 F 10.0  
14:00 248.2 F 14.9  
16:00 260.8 F 21.9  
18:00 272.1 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.4 F 29.8  
2:00 272.1 F 29.1  
3:00 271.8 F 29.5  
4:00 272.5 F 29.6  
  
Min 271.4 F 28.8  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 215.8 F 0.0  
10:00 226.7 F 0.0  
15:00 231.9 F 0.0  
20:00 228.4 F 0.0  
25:00 227.1 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 2:09  
HEATING: 18:24  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:48  
CYCLE COMPLETE

06/08/2023 01:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_