
Midmark M11 - v1.0.0
Total Cycles: 2510

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2022 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.4 F 0.0
2:00 184.7 F 3.1
4:00 184.1 F 7.5
6:00 206.5 F 12.8
8:00 222.8 F 7.8
10:00 236.8 F 13.1
12:00 238.2 F 10.4
14:00 235.5 F 8.5
16:00 250.7 F 15.5
18:00 262.7 F 22.4
20:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.2
2:00 272.1 F 28.6
3:00 271.2 F 28.2
4:00 272.2 F 29.0

Min 271.2 F 28.1
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 241.8 F 0.0
10:00 230.6 F 0.0
15:00 227.9 F 0.0
20:00 226.9 F 0.0
25:00 228.1 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:04
CYCLE COMPLETE

11/03/2022 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____