
Midmark M11 - v1.0.0
Total Cycles: 1157

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2021 01:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	109.6 F	0.0
2:00	121.3 F	0.0
4:00	152.3 F	0.0
6:00	188.5 F	0.0
8:00	208.7 F	0.1
10:00	212.6 F	0.3
12:00	213.4 F	0.4
14:00	213.9 F	0.5
16:00	214.5 F	0.7
18:00	215.3 F	0.9
20:00	216.2 F	1.2
22:00	217.6 F	1.7
24:00	219.6 F	2.3
26:00	222.9 F	3.4
28:00	229.5 F	5.8

C633: HEATUP MODE
PRESSURE LOW
CHAMBER TEMP 236F
CHAMBER PRESS 8.5PSI
** ITEMS NOT STERILE **

09/09/2021 01:47 PM