

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3631

Name: RITTER 1  
Sterilizer ID: V2178198  
08/10/2023 07:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 75.2 F 0.0  
4:00 89.1 F 0.0  
6:00 120.3 F 0.7  
8:00 151.8 F 3.3  
10:00 171.4 F 6.4  
12:00 188.2 F 9.7  
14:00 204.1 F 13.1  
16:00 213.5 F 5.0  
18:00 225.3 F 8.7  
20:00 234.7 F 12.7  
22:00 232.6 F 8.7  
24:00 242.2 F 11.9  
26:00 242.4 F 11.9  
28:00 254.3 F 17.9  
30:00 264.6 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.4 F 29.0  
2:00 271.8 F 29.1  
3:00 272.1 F 28.8  
4:00 272.2 F 29.4

Min 271.4 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
5:00 207.6 F 0.0  
10:00 220.8 F 0.0  
15:00 231.3 F 0.0  
20:00 226.2 F 0.0  
25:00 224.0 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 32:00  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:10:02  
CYCLE COMPLETE

08/10/2023 08:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_