
Midmark M11 - v1.0.5
Total Cycles: 1193

Name: RITTER2
Sterilizer ID: V2313228
06/16/2022 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.3 F 0.0
2:00 164.3 F 1.8
4:00 171.2 F 5.4
6:00 195.2 F 10.3
8:00 217.4 F 5.4
10:00 227.9 F 10.4
12:00 232.2 F 7.9
14:00 242.7 F 13.6
16:00 242.5 F 12.4
18:00 254.6 F 18.8
20:00 266.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.7
1:00 272.0 F 29.6
2:00 272.2 F 29.9
3:00 272.2 F 30.2
4:00 271.7 F 29.6

Min 271.6 F 29.5
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.9 F 0.0
5:00 216.0 F 0.0
10:00 228.9 F 0.0
15:00 209.4 F 0.0
20:00 209.6 F 0.0
25:00 207.7 F 0.0
30:00 203.6 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:32
CYCLE COMPLETE

06/16/2022 09:02 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____