

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 001498  
8:07 10/14/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:58  
PRESSURIZING 6:01  
STERILIZING 7:38  
136.5 C 220kPa 7:38  
136.4 C 222kPa 8:08  
136.3 C 221kPa 8:38  
136.0 C 219kPa 9:08  
136.4 C 217kPa 9:38  
136.3 C 220kPa 10:08  
135.8 C 219kPa 10:38  
136.3 C 220kPa 11:08  
VENTING 11:08  
AIR DRYING 12:12  
CYCLE COMPLETE 42:12

-----  
Digital Signature #  
2458863FE2C68AA0  
-----