

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4693

Name: RITTER 1  
Sterilizer ID: V2178198  
07/02/2024 05:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.7 F 0.0  
2:00 78.3 F 0.3  
4:00 88.0 F 1.6  
6:00 115.7 F 4.1  
8:00 146.3 F 7.2  
10:00 193.0 F 10.5  
12:00 212.1 F 2.8  
14:00 208.4 F 5.8  
16:00 218.5 F 9.4  
18:00 230.0 F 13.5  
20:00 233.4 F 9.0  
22:00 241.2 F 11.2  
24:00 242.9 F 11.9  
26:00 254.6 F 17.9  
28:00 265.3 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.4  
1:00 272.0 F 29.0  
2:00 272.3 F 28.8  
3:00 271.4 F 28.5  
4:00 272.7 F 29.2

Min 271.2 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
5:00 207.4 F 0.0  
10:00 216.9 F 0.0  
15:00 217.8 F 0.0  
20:00 230.5 F 0.0  
25:00 229.6 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 29:44  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:07:55  
CYCLE COMPLETE

07/02/2024 06:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_