
Midmark M11 - v1.0.5
Total Cycles: 2946

Name: RITTER2
Sterilizer ID: V2313228
09/28/2023 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 173.0 F 3.6
4:00 189.3 F 8.2
6:00 208.9 F 13.5
8:00 221.8 F 8.2
10:00 241.4 F 13.6
12:00 239.5 F 11.0
14:00 237.9 F 9.7
16:00 251.7 F 16.5
18:00 263.4 F 23.3
20:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.1 F 28.9
2:00 272.5 F 29.3
3:00 271.8 F 29.2
4:00 272.0 F 28.9

Min 271.4 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 251.6 F 0.0
10:00 234.7 F 0.0
15:00 230.8 F 0.0
20:00 221.4 F 0.0
25:00 217.1 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

09/28/2023 03:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____