
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001458
13:26 09/29/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:04
PRESSURIZING 5:37
STERILIZING 6:59
136.0 C 220kPa 6:59
136.0 C 219kPa 7:29
136.3 C 221kPa 7:59
136.2 C 220kPa 8:29
136.0 C 219kPa 8:59
136.5 C 223kPa 9:29
136.1 C 221kPa 9:59
136.4 C 223kPa 10:29
VENTING 10:29
AIR DRYING 11:33
CYCLE COMPLETE 41:33

Digital Signature #
8A943A4B66E40FFB
