

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3901

Name: RITTER2  
Sterilizer ID: V2313228  
08/28/2024 03:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.3 F 0.0  
2:00 191.6 F 2.7  
4:00 197.7 F 8.7  
6:00 219.2 F 6.9  
8:00 233.1 F 10.8  
10:00 233.9 F 8.5  
12:00 240.2 F 10.2  
14:00 248.7 F 14.7  
16:00 261.6 F 22.0  
18:00 272.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.0 F 28.6  
2:00 272.5 F 28.9  
3:00 272.9 F 29.3  
4:00 272.4 F 29.2  
  
Min 271.7 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
5:00 217.3 F 0.0  
10:00 228.0 F 0.0  
15:00 227.1 F 0.0  
20:00 224.3 F 0.0  
25:00 222.7 F 0.0  
30:00 219.5 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 18:14  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 00:57:28  
CYCLE COMPLETE

08/28/2024 04:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_