
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003910
8:47 29/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:22
STERILIZING 3:15
136.0 C 220kPa 3:15
135.9 C 221kPa 3:45
136.4 C 223kPa 4:15
136.2 C 218kPa 4:45
136.4 C 223kPa 5:15
136.1 C 220kPa 5:45
136.4 C 221kPa 6:15
136.4 C 220kPa 6:45
136.4 C 220kPa 6:45
VENTING 6:45
AIR DRYING 8:29
CYCLE COMPLETE 8:45

Digital Signature #
93A88DBFF6A0946C
