
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005864
9:25 29/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:58
STERILIZING 3:04
136.5 C 224kPa 3:04
136.2 C 221kPa 3:34
136.5 C 223kPa 4:04
136.4 C 222kPa 4:34
136.5 C 224kPa 5:04
136.4 C 223kPa 5:34
136.2 C 222kPa 6:04
136.4 C 223kPa 6:34
136.4 C 223kPa 6:34
VENTING 6:34
AIR DRYING 7:23
CYCLE COMPLETE 7:28

Digital Signature #
E01AA323D4B8D9F0
