
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000177
15:19 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:16
PRESSURIZING 4:16
STERILIZING 5:27
136.0 C 222kPa 5:27
136.4 C 226kPa 5:57
136.4 C 226kPa 6:27
136.3 C 225kPa 6:57
136.4 C 226kPa 7:27
136.4 C 227kPa 7:57
136.4 C 225kPa 8:27
136.3 C 225kPa 8:57
136.3 C 225kPa 8:57
VENTING 8:57
AIR DRYING 9:58
CYCLE COMPLETE 24:58

Digital Signature #
13DDDE25784244F3
