

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4856

Name: RITTER 1  
Sterilizer ID: V2178198  
08/27/2024 03:09 PM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.8 F 0.0  
2:00 127.2 F 0.2  
4:00 157.9 F 0.6  
6:00 189.5 F 2.9  
8:00 209.5 F 7.9  
10:00 224.7 F 13.3  
12:00 230.9 F 8.3  
14:00 241.5 F 11.8  
16:00 242.8 F 11.7  
18:00 240.1 F 10.2  
20:00 255.1 F 17.7  
22:00 255.3 F 17.8  
24:00 269.9 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.0  
1:00 271.7 F 28.0  
2:00 271.7 F 28.0  
3:00 271.9 F 27.9

Min 271.7 F 27.8  
Max 272.9 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 162.5 F 0.0  
5:00 199.2 F 0.0  
10:00 193.3 F 0.0  
15:00 196.5 F 0.0  
20:00 204.0 F 0.0  
25:00 204.7 F 0.0  
30:00 202.4 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 24:45  
STERILIZING: 3:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:01:27  
CYCLE COMPLETE

08/27/2024 04:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_