
Midmark M11 - v1.0.0
Total Cycles: 3479

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2023 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 184.2 F 3.5
4:00 202.1 F 8.0
6:00 214.9 F 13.0
8:00 225.2 F 8.7
10:00 236.2 F 10.8
12:00 241.7 F 11.7
14:00 239.5 F 10.2
16:00 253.3 F 17.1
18:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.0 F 29.0
2:00 272.3 F 28.8
3:00 271.4 F 28.4
4:00 272.5 F 29.3

Min 271.3 F 28.3
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 215.8 F 0.0
10:00 237.9 F 0.0
15:00 242.7 F 0.0
20:00 230.0 F 0.0
25:00 220.6 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

07/04/2023 12:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____