

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4937

Name: RITTER 1  
Sterilizer ID: V2178198  
09/25/2024 10:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.8 F 0.0  
2:00 148.1 F 0.0  
4:00 156.2 F 0.1  
6:00 180.6 F 0.4  
8:00 201.1 F 2.9  
10:00 214.3 F 6.5  
12:00 226.6 F 10.4  
14:00 230.8 F 7.9  
16:00 237.3 F 11.0  
18:00 234.4 F 8.2  
20:00 242.8 F 11.4  
22:00 246.1 F 13.1  
24:00 257.7 F 19.2  
26:00 267.9 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.4 F 28.8  
2:00 272.1 F 28.4  
3:00 271.4 F 28.0  
4:00 272.1 F 28.7

Min 271.3 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.1 F 0.0  
5:00 195.6 F 0.0  
10:00 190.2 F 0.0  
15:00 189.1 F 0.0  
20:00 203.2 F 0.0  
25:00 216.3 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 27:29  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:05:59  
CYCLE COMPLETE

09/25/2024 11:08 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_