

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4957

Name: RITTER 1  
Sterilizer ID: V2178198  
10/03/2024 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 72.3 F 0.0  
4:00 89.1 F 0.0  
6:00 125.2 F 0.1  
8:00 158.8 F 0.3  
10:00 183.4 F 1.8  
12:00 205.5 F 5.1  
14:00 221.0 F 8.1  
16:00 230.3 F 11.5  
18:00 224.9 F 4.9  
20:00 235.9 F 9.9  
22:00 237.6 F 9.2  
24:00 243.7 F 12.0  
26:00 239.9 F 10.1  
28:00 252.8 F 16.4  
30:00 263.7 F 22.8  
32:00 272.1 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.4 F 28.6  
2:00 271.4 F 27.9  
3:00 272.5 F 28.7  
4:00 271.7 F 28.0  
  
Min 271.2 F 27.8  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
5:00 227.1 F 0.0  
10:00 179.2 F 0.0  
15:00 172.6 F 0.0  
20:00 172.9 F 0.0  
25:00 172.7 F 0.0  
30:00 178.2 F 0.0  
  
Drying Complete

FILLING: 0:44  
HEATING: 32:12  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:10:19  
CYCLE COMPLETE

10/03/2024 09:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_