
Midmark M11 - v1.0.5
Total Cycles: 2258

Name: RITTER2
Sterilizer ID: V2313228
04/18/2023 06:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.7 F 0.0
2:00 78.8 F 0.0
4:00 122.9 F 0.0
6:00 116.3 F 0.0
8:00 167.2 F 1.3
10:00 189.9 F 4.5
12:00 208.7 F 8.3
14:00 225.2 F 12.2
16:00 222.9 F 6.3
18:00 235.5 F 10.6
20:00 234.4 F 8.2
22:00 242.2 F 12.4
24:00 240.3 F 10.9
26:00 252.2 F 17.0
28:00 262.8 F 23.4
30:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.6 F 29.6
2:00 272.1 F 29.8
3:00 272.1 F 29.2
4:00 272.5 F 30.0

Min 271.7 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 209.7 F 0.0
10:00 207.5 F 0.0
15:00 226.3 F 0.0
20:00 218.1 F 0.0
25:00 220.0 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:20
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:08:20
CYCLE COMPLETE

04/18/2023 08:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____