

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1809

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2022 09:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.2 F 0.0  
2:00 148.2 F 0.0  
4:00 151.2 F 0.0  
6:00 183.9 F 2.8  
8:00 204.2 F 7.1  
10:00 221.7 F 12.5  
12:00 228.9 F 8.3  
14:00 240.3 F 12.5  
16:00 241.0 F 11.7  
18:00 239.9 F 10.6  
20:00 253.3 F 17.4  
22:00 264.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.3 F 29.1  
2:00 272.3 F 29.6  
3:00 271.7 F 28.7  
4:00 272.7 F 29.4

Min 271.4 F 28.7  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.8 F 0.0  
5:00 229.6 F 0.0  
10:00 221.9 F 0.0  
15:00 224.4 F 0.0  
20:00 220.7 F 0.0  
25:00 224.1 F 0.0  
30:00 219.2 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 23:48  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:01:51  
CYCLE COMPLETE

12/06/2022 10:14 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_