
Midmark M11 - v1.0.5
Total Cycles: 3761

Name: RITTER2
Sterilizer ID: V2313228
07/09/2024 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 72.1 F 0.0
4:00 95.6 F 0.0
6:00 102.7 F 0.1
8:00 139.5 F 1.8
10:00 165.5 F 4.5
12:00 186.8 F 7.6
14:00 200.4 F 11.0
16:00 214.2 F 4.8
18:00 222.5 F 8.1
20:00 232.4 F 12.3
22:00 233.2 F 8.7
24:00 241.9 F 11.8
26:00 243.5 F 12.4
28:00 254.9 F 18.3
30:00 265.1 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.6
2:00 271.5 F 28.9
3:00 272.8 F 29.7
4:00 271.8 F 28.9

Min 271.4 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 206.7 F 0.0
10:00 205.6 F 0.0
15:00 193.9 F 0.0
20:00 196.9 F 0.0
25:00 197.0 F 0.0
30:00 198.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:59
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:10
CYCLE COMPLETE

07/09/2024 09:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____