
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002340
8:49 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:06
STERILIZING 4:18
136.0 C 220kPa 4:18
136.2 C 222kPa 4:48
136.4 C 223kPa 5:18
136.4 C 224kPa 5:48
136.4 C 223kPa 6:18
136.0 C 222kPa 6:48
136.4 C 224kPa 7:18
136.2 C 222kPa 7:48
136.2 C 222kPa 7:48
VENTING 7:48
AIR DRYING 8:40
CYCLE COMPLETE 32:02

Digital Signature #
60B45DCCDAA5E7BE
