
Midmark M11 - v1.0.5
Total Cycles: 3150

Name: RITTER2
Sterilizer ID: V2313228
11/21/2023 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 171.9 F 2.3
4:00 187.4 F 6.9
6:00 211.0 F 12.5
8:00 223.2 F 7.7
10:00 237.7 F 12.8
12:00 240.6 F 11.7
14:00 240.9 F 11.0
16:00 255.0 F 18.3
18:00 267.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.6 F 29.2
2:00 271.9 F 29.3
3:00 272.1 F 29.0
4:00 272.1 F 29.5

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.7 F 0.0
5:00 216.2 F 0.0
10:00 204.8 F 0.0
15:00 201.5 F 0.0
20:00 200.6 F 0.0
25:00 198.7 F 0.0
30:00 197.9 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

11/21/2023 12:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____