
Midmark M11 - v1.0.0
Total Cycles: 4323

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2024 02:59 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.8 F 0.0
2:00 108.5 F 1.2
4:00 127.2 F 3.5
6:00 155.9 F 6.7
8:00 179.3 F 10.3
10:00 211.5 F 5.6
12:00 213.7 F 7.2
14:00 226.0 F 11.4
16:00 228.2 F 7.1
18:00 240.3 F 12.4
20:00 238.5 F 10.2
22:00 250.3 F 16.1
24:00 261.4 F 22.3
26:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.2 F 29.1
2:00 272.4 F 29.5
3:00 272.0 F 29.3
4:00 272.1 F 29.1

Min 271.6 F 28.9
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 228.6 F 0.0
10:00 225.7 F 0.0
15:00 226.4 F 0.0
20:00 229.1 F 0.0
25:00 228.4 F 0.0
30:00 227.8 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 26:35
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:08
CYCLE COMPLETE

02/22/2024 04:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____