
Midmark M11 - v1.0.0
Total Cycles: 4779

Name: RITTER 1
Sterilizer ID: V2178198
07/31/2024 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.0 F 0.0
2:00 136.5 F 1.2
4:00 164.2 F 4.4
6:00 184.0 F 8.4
8:00 202.7 F 13.2
10:00 219.3 F 5.2
12:00 230.5 F 9.3
14:00 236.4 F 10.9
16:00 239.8 F 10.8
18:00 235.8 F 8.5
20:00 250.0 F 15.2
22:00 261.7 F 21.8
24:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.1
1:00 272.5 F 28.6
2:00 272.5 F 28.9
3:00 271.5 F 28.0
4:00 272.4 F 28.5

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.9 F 0.0
5:00 194.6 F 0.0
10:00 196.2 F 0.0
15:00 194.3 F 0.0
20:00 201.9 F 0.0
25:00 201.2 F 0.0
30:00 235.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:13
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

07/31/2024 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____