

-----  
Mississauga, ON  
Canada  
-----

STATIM 5000 S501R751  
SN 511020J00004  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000796  
14:59 16/03/2022  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:48  
PRESSURIZING 5:32  
STERILIZING 6:54  
136.0 C 220kPa 6:54  
136.3 C 223kPa 7:24  
136.4 C 221kPa 7:54  
136.0 C 219kPa 8:24  
136.2 C 222kPa 8:54  
136.4 C 221kPa 9:24  
136.3 C 223kPa 9:54  
136.4 C 223kPa 10:24  
136.4 C 223kPa 10:24  
VENTING 10:24  
AIR DRYING 11:24  
CYCLE COMPLETE 23:22  
-----

Digital Signature #  
7495CA1743593EAE  
-----