

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2584

Name: RITTER 1  
Sterilizer ID: V2178198  
11/21/2022 11:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
2:00 189.3 F 4.4  
4:00 196.7 F 9.2  
6:00 212.6 F 7.6  
8:00 228.0 F 9.0  
10:00 235.9 F 10.0  
12:00 240.6 F 11.5  
14:00 238.6 F 9.7  
16:00 252.7 F 16.6  
18:00 264.8 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.2  
1:00 272.3 F 28.9  
2:00 271.5 F 28.2  
3:00 271.6 F 28.7  
4:00 272.2 F 28.8  
  
Min 271.0 F 28.0  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 233.4 F 0.0  
10:00 223.8 F 0.0  
15:00 226.6 F 0.0  
20:00 231.5 F 0.0  
25:00 229.8 F 0.0  
30:00 230.0 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 19:51  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

11/21/2022 12:21 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_