
Midmark M11 - v1.0.0
Total Cycles: 4089

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2023 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.5 F 0.0
2:00 119.8 F 0.7
4:00 119.8 F 2.6
6:00 141.7 F 5.3
8:00 166.8 F 8.3
10:00 187.0 F 12.0
12:00 206.8 F 4.1
14:00 214.2 F 8.3
16:00 226.5 F 12.3
18:00 229.5 F 8.4
20:00 243.7 F 13.4
22:00 241.5 F 11.4
24:00 253.3 F 17.1
26:00 263.5 F 23.2
28:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 271.9 F 29.2
2:00 272.5 F 29.2
3:00 272.1 F 29.4
4:00 272.5 F 29.2

Min 270.7 F 28.3
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 210.0 F 0.0
10:00 200.7 F 0.0
15:00 193.9 F 0.0
20:00 191.8 F 0.0
25:00 189.5 F 0.0
30:00 191.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:11
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:51
CYCLE COMPLETE

12/06/2023 11:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____