
Midmark M11 - v1.0.0
Total Cycles: 951

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2021 06:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	73.5 F	0.0
2:00	97.4 F	0.0
4:00	92.4 F	0.0
6:00	128.0 F	0.0
8:00	156.1 F	0.2
10:00	176.3 F	0.2
12:00	195.4 F	1.3
14:00	213.1 F	4.0
16:00	222.5 F	6.4
18:00	228.4 F	9.5
20:00	238.1 F	13.0
22:00	234.1 F	8.4
24:00	243.9 F	12.9
26:00	240.7 F	10.4
28:00	237.6 F	8.7
30:00	249.3 F	14.5
32:00	259.8 F	20.2
34:00	269.4 F	26.3

STERILIZING

mm:ss	Degrees	PSI
0:00	271.8 F	27.8
1:00	271.9 F	28.2
2:00	272.5 F	28.3
3:00	271.6 F	27.7
4:00	272.2 F	28.4

Min 271.5 F 27.6
Max 272.7 F 28.6

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	217.9 F	0.0
5:00	203.2 F	0.0
10:00	195.0 F	0.0
15:00	196.0 F	0.0
20:00	194.9 F	0.0
25:00	237.0 F	0.0
30:00	236.7 F	0.0

Drying Complete

FILLING: 0:51
HEATING: 35:00
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:13:14
CYCLE COMPLETE

06/29/2021 08:07 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____