

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3579

Name: RITTER 1  
Sterilizer ID: V2178198  
07/25/2023 12:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.3 F 0.0  
2:00 183.2 F 2.1  
4:00 206.2 F 5.9  
6:00 209.3 F 10.3  
8:00 219.2 F 6.7  
10:00 228.7 F 8.4  
12:00 239.1 F 13.0  
14:00 238.7 F 10.3  
16:00 237.6 F 8.9  
18:00 249.1 F 14.7  
20:00 259.9 F 20.7  
22:00 270.2 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.2  
1:00 272.6 F 29.0  
2:00 271.6 F 28.2  
3:00 272.5 F 29.0  
4:00 271.9 F 28.4

Min 271.3 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 212.2 F 0.0  
10:00 214.3 F 0.0  
15:00 216.8 F 0.0  
20:00 217.4 F 0.0  
25:00 219.4 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 22:48  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:02:12  
CYCLE COMPLETE

07/25/2023 01:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_