
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002560
10:21 05/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:47
STERILIZING 3:53
136.0 C 220kPa 3:53
135.9 C 221kPa 4:23
136.1 C 223kPa 4:53
136.4 C 226kPa 5:23
136.3 C 223kPa 5:53
136.1 C 222kPa 6:23
136.3 C 224kPa 6:53
136.2 C 223kPa 7:23
136.2 C 223kPa 7:23
VENTING 7:23
AIR DRYING 8:15
CYCLE COMPLETE 38:15

Digital Signature #
FF5ED72C600F562A
